UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/717,798	11/19/2003	Amir Abolfathi	9063.ALG.P	8591
56188 7590 03/29/2007 GREENBERG TRAURIG, LLP (SV2)			EXAMINER	
2450 Colorado Avenue, Suite 400E			O'CONNOR, CARY E	
Santa Monica, CA 90404			ART UNIT	PAPER NUMBER
			3732	
SHORTENED STATUTOR	Y PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE	
3 MONTHS		03/29/2007	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Paper No(s)/Mail Date \_

2) Notice of Draftsperson's Patent Drawing Review (PTO-948)

3) Information Disclosure Statement(s) (PTO/SB/08)

Paper No(s)/Mail Date. \_

6) \_\_ Other: \_\_

Notice of Informal Patent Application

Application/Control Number: 10/717,798

Art Unit: 3732

#### **DETAILED ACTION**

### Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 15-19 and 51 are rejected under 35 U.S.C. 102(e) as being anticipated by Kaza (3002/0129565). Kaza shows a dental tray (paragraph 0060) for creating dental impressions. Kaza does not specify that the tray contains radiopaque material. However the claims recite that one of the tray material and the impression material has the radiopaque agent. Accordingly, the in the case that the impression material has the agent, Kaza anticipates the claims because the impression material is not positively claimed, leaving the claim limited to a dental tray made of a material, the radiation source, etc. The recitation that the respective radiopacity of the impression material and the tray material approximately match each other has not been given patentable weight in the claim because the impression material is not positively claimed. The system of Kaza includes a system for creating a digital model of a patient's teeth comprising radiation source (an X-ray source) 802, a scintillator 812, a radiation detector coupled to the scintillator, a rotatable table 804 positioned between the radiation source and the scintillator, and a computer 822 coupled to the detector.

Application/Control Number: 10/717,798

Art Unit: 3732

## Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claim 20 is rejected under 35 U.S.C. 103(a) as being unpatentable over Kaza (2003/0129565) in view of in view of Coscina (3,878,610). Kaza is silent as to the form of the tray. Coscina shows a dental tray 10 comprising a base 26 having a plurality of prongs, a first wall 28 extending from one side of the base, at least one tearable portion formed on one end of one prong, the detachable portion being removable to shorten the prong length (see column 7, lines 36+). It would have been obvious to one of ordinary skill in the art at the time the invention was made to provide system of Kaza with the tray of Coscina, in order to contain the impression material and provide a tray that is adjustable.

# Allowable Subject Matter

Claims 1-14, 21-50 are allowed.

# Response to Arguments

Applicant's arguments with respect to claims 15 and 51 have been considered but are most in view of the new ground(s) of rejection.

Application/Control Number: 10/717,798

Art Unit: 3732

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Cary E. O'Connor whose telephone number is 571-272-4715. The examiner can normally be reached on M-Th 7:00-3:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Cris Rodriguez can be reached on 571-2724964. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Cary E. O'Connor Primary Examiner Art Unit 3732